

LAMPIRAN

Lampiran 1. Surat Keterangan Layak Etik



Kementerian Kesehatan

Poltekkes Yogyakarta

Komite Etik Penelitian Kesehatan

Jalan Tata Bumi No. 3, Banyuraden, Gamping,
Sleman, D.I. Yogyakarta 55293

(0274) 617601

<https://poltekkesjogja.ac.id>

KETERANGAN LAYAK ETIK DESCRIPTION OF ETHICAL EXEMPTION "ETHICAL EXEMPTION"

No.DP.04.03/e-KEPK.1/437/2025

Protokol penelitian versi 1 yang diusulkan oleh :
The research protocol proposed by

Peneliti utama : Fadilah Yuliana Rahmawati
Principal In Investigator

Nama Institusi : Poltekkes Kemenkes Yogyakarta
Name of the Institution

Dengan judul:
Title

"Pengaruh Penyimpanan Serum Kontrol Liofilisat Komersial Setelah Rekonstitusi Yang Disimpan Pada Suhu -20°C Terhadap Hasil Pemeriksaan Kadar Ureum"

"Effect of Storage of Commercial Lyophilizate Control Serum After Reconstitution Stored at -20°C on the Results of Ureum Level Examination"

Dinyatakan layak etik sesuai 7 (tujuh) Standar WHO 2011, yaitu 1) Nilai Sosial, 2) Nilai Ilmiah, 3) Pemerataan Beban dan Manfaat, 4) Risiko, 5) Bujukan/Eksploitasi, 6) Kerahasiaan dan Privacy, dan 7) Persetujuan Setelah Penjelasan, yang merujuk pada Pedoman CIOMS 2016. Hal ini seperti yang ditunjukkan oleh terpenuhinya indikator setiap standar.

Declared to be ethically appropriate in accordance to 7 (seven) WHO 2011 Standards, 1) Social Values, 2) Scientific Values, 3) Equitable Assessment and Benefits, 4) Risks, 5) Persuasion/Exploitation, 6) Confidentiality and Privacy, and 7) Informed Consent, referring to the 2016 CIOMS Guidelines. This is as indicated by the fulfillment of the indicators of each standard.

Pernyataan Laik Etik ini berlaku selama kurun waktu tanggal 17 Maret 2025 sampai dengan tanggal 17 Maret 2026.

This declaration of ethics applies during the period March 17, 2025 until March 17, 2026.



March 17, 2025
Chairperson,



Dr. drg. Wiworo Haryani, M.Kes.

Lampiran 2. Surat Izin Peminjaman Tempat Penelitian



Kementerian Kesehatan
Poltekkes Yogyakarta
Jalan Tata Bumi No. 3, Banyuraden, Gamping,
Sleman, D.I. Yogyakarta 55293
(0274) 617601
<https://poltekkesjogja.ac.id>

Kepada
Yth. **Ketua Jurusan Teknologi Laboratorium Medis**
Poltekkes Kemenkes Yogyakarta
Di tempat

Dengan Hormat,

Sehubungan dengan Penelitian Skripsi oleh Mahasiswa Sarjana Terapan Jurusan Teknologi Laboratorium Medis Poltekkes Kemenkes Yogyakarta, saya selaku peneliti:


Nama : Fadilah Yuliana Rahmawati
NIM : P71342324090
No. HP : 082196966907
Judul : Pengaruh Penyimpanan Serum Kontrol Liofilisat Komersial Setelah Rekonstitusi Yang Disimpan Pada Suhu -20°C Terhadap Hasil Pemeriksaan Kadar Ureum

Memohon izin untuk meminjam Ruang Laboratorium Kimia Klinik untuk melakukan Penelitian Skripsi. Adapun kegiatan yang akan dilaksanakan pada:


Hari, tanggal : Senin-Sabtu, (10 Maret – 22 April 2025)
Waktu : 08.00 – 17.00
Tempat : Laboratorium Kimia Klinik Jurusan Teknologi Laboratorium Medis Poltekkes Kemenkes Yogyakarta

Demikian surat permohonan peminjaman ini kami sampaikan. Adapun daftar alat dan bahan terlampir. Atas perhatian dan izin yang diberikan, kami ucapkan terimakasih.

Yogyakarta, 7 Maret 2025

Penanggungjawab Laboratorium Kimia Klinik	Pembimbing Tugas Akhir	Peneliti
		
Subrata Tri Widada, SKM, M.Sc. NIP. 19631128 198303 1 001	M. Atik Martsiningsih, S.Si., M.Sc NIP. 19680323 198803 2 002	Fadilah Yuliana Rahmawati NIM P71342324090

Mengetahui,
PJ Laboratorium Jurusan
Teknologi Laboratorium Medis


Zulfikar Husni Faruq, M.Si.
NIP. 19890725 201902 1 001

Tembusan :

1. Mahasiswa
2. Koordinator Laboratorium Jurusan Teknologi Laboratorium Medis
3. Penanggung Jawab Ruang Laboratorium Jurusan Teknologi Laboratorium Medis
4. Petugas Laboratorium Jurusan Teknologi Laboratorium Medis
5. Satpam Jurusan Teknologi Laboratorium Medis

 Dipindai dengan CamScanner

Lampiran 3. Surat Telah Menyelesaikan Penelitian



Kementerian Kesehatan
Poltekkes Yogyakarta

Jalan Tata Bumi No. 3, Banyuraden, Gamping,
Sleman, D.I. Yogyakarta 55293
(0274) 617601
<https://poltekkesjogja.ac.id>

SURAT KETERANGAN
Nomor : TL.02.01.470

Dengan ini menyatakan bahwa :

Nama : Fadilah Yuliana Rahmawati
NIM : P71342324090
Institusi : Prodi Sarjana Terapan Teknologi Laboratorium Medik Poltekkes
Kemenkes Yogyakarta.
Judul penelitian : Pengaruh Penyimpanan Serum Kontrol Liofilisat Komersial Setelah
Rekonstitusi Yang Disimpan Pada Suhu -20°C Terhadap Hasil
Pemeriksaan Kadar Ureum

Bahwasanya mahasiswa tersebut di atas telah selesai melakukan penelitian di Laboratorium
Kimia Klinik Jurusan Teknologi Laboratorium Medis Poltekkes Kemenkes Yogyakarta.

Demikian surat keterangan ini dibuat untuk dapat digunakan sebagaimana mestinya.

Yogyakarta, 30 April 2025
Ketua Jurusan

Muji Rahayu, S.Si Apt. M.Sc.
NIP. 196606151985112001

Tembusan :

1. Mahasiswa
2. Koordinator Laboratorium Jurusan Teknologi Laboratorium Medis
3. Penanggung Jawab Ruang Laboratorium Jurusan Teknologi Laboratorium Medis
4. Petugas Laboratorium Jurusan Teknologi Laboratorium Medis
5. Satpam Jurusan Teknologi Laboratorium Medis


Lampiran 4. Instrumen Uji Pendahuluan

-
1. Dimana tempat anda bekerja atau pernah bekerja
 - a. Rumah Sakit
 - b. Puskesmas
 - c. Klinik
 2. Alat pemeriksaan kimia klinik apa yang anda gunakan*
 - a. Kimia Analyze
 - b. Spektrofotometer
 3. Jenis serum kontrol apa yang digunakan
 - a. Assayed (memiliki nilai rujukan)
 - b. Unassayed (tidak memiliki nilai rujukan)
 - c. Pooled sera / Kumpulan serum
 4. Bentuk bahan kontrol apa yang anda gunakan
 - a. Cair
 - b. Serbuk
 5. Bagaimana prosedur yang benar dalam mengelolah serum kontrol yang anda gunakan, apakah langsung digunakan di botol atau dipindahkan kedalam per cup
 - a. Dipindahkan kedalam per cup
 - b. Langsung digunakan
 6. Dimana anda menyimpan serum kontrol yang anda gunakan
 - a. Pada suhu - 20 C di bagian *Freezer* (Kulkas bagian atas)
 - b. Pada suhu 2-8 C di bagian *Chiller* (Kulkas bagian bawah)
 7. Bagaimana anda melakukan quality control di tempat anda bekerja
 - a. Dilakukan setiap hari
 - b. Dilakukan hanya saat ada permintaan pemeriksaan
 8. Seberapa sering permintaan pemeriksaan Ureum di tempat anda bekerja*
 - a. Rutin (Tiap hari selalu dilakukan)
 - b. Sering (lebih dari 10-20 kali dalam sebulan)
 - c. Jarang (dibawah 10 kali dalam sebulan)
-

Lampiran 5. Diagram Hasil Uji Pendahuluan


Dimana tempat anda bekerja atau pernah bekerja

121 jawaban




Alat pemeriksaan kimia klinik apa yang anda gunakan

121 jawaban




Jenis serum kontrol apa yang digunakan

121 jawaban




Bentuk bahan kontrol apa yang anda gunakan

121 jawaban




Bagaimana prosedur yang benar dalam mengelolah serum kontrol yang anda gunakan, apakah langsung digunakan di botol atau dipindahkan kedalam per cup

121 jawaban




Dimana anda menyimpan serum kontrol yang anda gunakan

121 jawaban




Bagaimana anda melakukan quality control di tempat anda bekerja

121 jawaban



Seberapa sering permintaan pemeriksaan Ureum di tempat anda bekerja

121 jawaban



Lampiran 6. Insert Kit Serum Control

HUMAN MULTISERA NORMAL						LOT #19309	EXP: 2027-10-30
COMPONENTE COMPONENT	VALOR TARGET	RANGE	1 SD	2 SD	UNIDAD UNIT	METODO METHOD	
ALP	277	205-319	211	422		Dehydrogenase buffer DEA, 37°C	
	216	184-249	164	328	U/L	Dehydrogenase buffer DEA, 30°C	
	178	151-205	135	27	U/L	Dehydrogenase buffer DEA, 25°C	
ALT/AST	37.9	30.3-45.6	3.39	7.58		Tiss no PSP IFCC/Choc, 37°C	
	28	22.4-33.6	2.8	5.6	U/L	Tiss no PSP IFCC/Choc, 30°C	
	21.4	17.1-25.7	2.14	4.28	U/L	Tiss no PSP IFCC/Choc, 25°C	
AST/AST	34.8	27.8-42	3.98	7.92		Tiss no PSP IFCC/Choc, 37°C	
	23.7	19.3-28.1	2.37	4.74	U/L	Tiss no PSP IFCC/Choc, 30°C	
	16.8	13.3-20.2	1.71	3.42	U/L	Tiss no PSP IFCC/Choc, 25°C	
Albumin	4.15	3.51-4.79	0.32	0.64	g/dL	Bromocresol green	
Albumin	4.15	35.2-47.8	3.15	6.3	g/L	Bromocresol green	
Amylase	90	76.5-104	6.75	13.5	U/L	Total T Liquid stable pH9.0, 37°C	
Amylase	0.78	0.61-0.97	0.09	0.18	mg/dL	DPD	
	13.3	10.6-15.4	1.46	2.92	mg/dL	DPD	
Bilirubin Direct	0.53	0.41-0.65	0.06	0.12	mg/dL	With sample blank Modified J&G	
Bilirubin Direct	9.06	7.08-11	0.99	1.98	mg/dL	With sample blank Modified J&G	
	0.81	0.63-0.99	0.09	0.18	mg/dL	Without sample blank Modified J&G	
	13.8	10.8-16.9	1.31	3.02	mg/dL	Without sample blank Modified J&G	
Bilirubin Total	1.34	1.08-1.6	0.13	0.26	mg/dL	DPD	
	22.9	18.3-27.5	2.29	4.58	mg/dL	DPD	
Bilirubin Total	1.6	1.26-1.94	0.17	0.34	mg/dL	With sample blank Modified J&G	
	27.4	21.5-33.2	2.53	5.89	mg/dL	With sample blank Modified J&G	
	1.83	1.49-1.97	0.17	0.34	mg/dL	Without sample blank Modified J&G	
	27.8	21.9-33.8	2.58	5.95	mg/dL	Without sample blank Modified J&G	
CK-Tot	206	171-247	16.8	37.6	U/L	CK-MAC substrate start (D5K), 37°C	
	131	107-154	11.8	23.6	U/L	CK-MAC substrate start (D5K), 30°C	
	89.2	73.1-103	8.03	16.1	U/L	CK-MAC substrate start (D5K), 25°C	
CK-MAC	8.04	7.24-8.84	0.4	0.8	mg/dL	Aspartate III	
	2.01	1.81-2.21	0.1	0.2	mg/dL	Aspartate III	
	7.71	6.93-8.46	0.39	0.78	mg/dL	Aspartate III	
	1.93	1.73-2.13	0.1	0.2	mg/dL	Aspartate III	
CK-MAC	80.5	72.8-88.2	3.96	7.72	mg/dL	Aspartate III	
CK-MAC	142	123-161	9.37	18.7	mg/dL	Aspartate III	
CK-MAC	3.88	3.2-4.6	0.24	0.48	mg/dL	Aspartate III	
CK-MAC	54.55	43.63-65.47	5.45	10.92	U/L	Aspartate III	
CK-MAC	1.22	0.98-1.46	0.12	0.24	mg/dL	Aspartate III	
CK-MAC	10.6	8.52-12.9	1.06	2.16	mg/dL	Aspartate III	

HUMAN MULTISERUM NORMAL						LOT #19309	EXP: 2027-10-30
COMPONENTE	VALOR	RANGE	1 SD	2 SD	UNIDAD	METODO	
GGT	44.8	38.1-51.5	3.30	6.72	U/L	5-glutamyl-L-α-ketooctonamide 37°C	
	35.2	29.0-40.5	2.64	5.28	U/L	5-glutamyl-L-α-ketooctonamide 30°C	
	28	23.8-32.2	2.1	4.2	U/L	5-glutamyl-L-α-ketooctonamide 25°C	
Glucose	97	82.5-112	7.27	14.5	mg/dL	Glucose Oxidase	
Glucose	5.38	4.58-6.18	0.4	0.9	mg/dL	Glucose Oxidase	
Iron	114	80.4-135	10.3	20.6	mg/dL	Cerium sulfate	
	20.4	16.7-24.1	1.64	3.28	mg/dL	Cerium sulfate	
Iron	94	77.1-111	8.46	16.9	mg/dL	Cerium sulfate	
	16.8	13.8-19.8	1.51	3.02	mg/dL	Cerium sulfate	
LDH	306	237-406	29.7	59.4	U/L	Phosphate → Lactate SPBC, 37°C	
	286	249-329	21.5	43	U/L	Phosphate → Lactate SPBC, 30°C	
	201	171-231	15.1	30.2	U/L	Phosphate → Lactate SPBC, 25°C	
Lipase	43.2	34.6-51.8	4.32	8.64	U/L	Enzymatic colorimetric, 37°C	
Lipase	2.27	1.99-2.55	0.14	0.28	mg/dL	Calmagite BR	
Magnesium	0.84	0.82-1.06	0.06	0.12	mg/dL	Calmagite BR	
Magnesium	2.16	1.9-2.42	0.13	0.26	mg/dL	Calmagite BR	
	0.89	0.77-1.01	0.09	0.12	mg/dL	Calmagite BR	
Phosphorus	4.59	3.9-5.26	0.34	0.68	mg/dL	Phosphomolybdate UV	
Phosphorus	1.48	1.26-1.7	0.11	0.22	mg/dL	Phosphomolybdate UV	
Phosphorus	3.97	3.57-4.37	0.2	0.4	mg/dL	ISE direct	
Potassium	3.97	3.57-4.37	0.2	0.4	mg/dL	ISE indirect	
Potassium	4.12	3.7-4.54	0.21	0.42	mmol/L	ISE indirect	
Potassium	140	126-154	7	14	mmol/L	ISE direct	
Sodium	140	126-154	7	14	mmol/L	ISE indirect	
Sodium	138	124-152	6.9	13.8	mmol/L	ISE indirect	
TIBC	224	177-271	23.7	47.4	g/dL	Saturation	
TIBC	40.1	31.6-48.6	4.25	8.5	g/dL	Saturation	
Total Protein	5.58	4.46-6.7	0.56	1.12	g/L	Buret endpoint	
Total Protein	5.8	4.6-6.7	0.58	1.12	g/L	Buret endpoint	
Triglycerides	91.2	76.4-106	7.39	14.8	mg/dL	Lipase/GOD-PAP no correction	
Triglycerides	103	87-119	8.08	16.16	mg/dL	Lipase/GOD-PAP no correction	
Urea/Nitrogen	42.6	36.2-49	3.19	6.38	mg/dL	UV Enzymatic	
Urea/Nitrogen	7.09	6.03-8.15	0.53	1.05	mg/dL	UV Enzymatic	
Urea/Nitrogen	6.0	5.2-6.8	0.4	0.8	mg/dL	UV Enzymatic	
Urea/Nitrogen	357	310-404	23.9	47.2	mg/dL	UV Enzymatic	

Lampiran 7. Insert Kit Reagen Glory Ureum



UREA Berthelot




GD-URB100 2 x 50 mL CONTENTS R1. Reagent 1 x 2 mL R2. Reagent 1 x 48 mL R3. Reagent 1 x 50 mL CAL. Standard 1 x 3 mL	GD-URB400 4 x 100 mL CONTENTS R1. Reagent 2 x 4 mL R2. Reagent 2 x 96 mL R3. Reagent 2 x 100 mL CAL. Standard 1 x 3 mL	UREA Urease/Salicylate Enzymatic colorimetric method ENDPOINT
--	--	---

For in vitro diagnostic use only

PRINCIPLE

Urea is hydrolyzed by urease^{1,2} into ammonia and carbon dioxide. The ammonia generated reacts with alkaline hypochlorite and sodium salicylate in presence of sodium nitroprusside as coupling agent to yield a green chromophore. The intensity of the color formed is proportional to the concentration of urea in the sample.



REAGENT COMPOSITION

- R1** Enzyme reagent. Urease > 500 U/mL. Stabilizers.
- R2** Buffered chromogen. Phosphate buffer 20 mmol/L pH 6.9, EDTA 2 mmol/L, sodium salicylate 60 mmol/L, sodium nitroprusside 3.4 mmol/L.
- R3** Alkaline hypochlorite. Sodium hypochlorite 10 mmol/L, NaOH 150 mmol/L.
- CAL** Urea standard. Urea 50 mg/dL (8.3 mmol/L) Organic matrix based primary standard. Concentration value is traceable to Standard Reference Material 909b.

STORAGE AND STABILITY

Store at 2-8°C.
All the kit compounds are stable until the expiry date stated on the label. Do not use reagents over the expiration date.
Store the vials tightly closed, protected from light and prevented contaminations during the use.

Discard if appear signs of deterioration:

- Presence of particles and turbidity.
- Blank absorbance (A) at 600 nm > 0.110 in 1cm cuvette.

REAGENT PREPARATION

Working reagent. Mix 1 volume of R1 + 24 volumes of R2. Stable for 4 weeks at 2-8°C and for 7 days at 15-25°C.

SAMPLES

Serum or heparinized plasma free of hemolysis and urine (see Notes). Other anticoagulants (ammonium heparinate or double oxalate of potassium and ammonium) must not be used.
Urea in serum, plasma or urine is stable 7 days at 2-8°C. Freeze for longer storage.

INTERFERENCES

- Lipemia (intralipid 20 g/L) does not interfere.
- Bilirubin (40 mg/dL) does not interfere
- Hemoglobin (>2 g/L) may affect the results.
- Other drugs and substances may interfere³.

MATERIALS REQUIRED

- Photometer or colorimeter capable of measuring absorbances at 600 ± 10 nm.
- Constant temperature incubator set at 37°C.
- Cuvettes with 1-cm pathlength.
- Pipettes to measure reagent and samples.

PROCEDURE

1. Bring reagents and samples to room temperature.
2. Pipette into a cuvette:

TUBES	Blank	Sample	CAL. Standard
Working reagent	1.0 mL	1.0 mL	1.0 mL
Sample	-	10 µL	-
CAL. Standard	-	-	10 µL

3. Mix and incubate for 5 minutes at 37°C or for 10 minutes at room temperature (16-25°C).
4. Pipette:

R3	1.0 mL	1.0 mL	1.0 mL
----	--------	--------	--------

5. Mix thoroughly and incubate the tubes for 5 minutes at 37°C or for 10 minutes at room temperature (16-25°C).
5. Read the absorbance (A) of the samples and the standard at 600 nm against the reagent blank.

The color is stable for at least 2 hours protected from light.

CALCULATIONS

Serum, plasma

$$A_{\text{Sample}} \times C_{\text{Standard}} = \text{mg/dL urea}$$

A Standard

Samples with concentrations higher than 300 mg/dL (50 mmol/L) should be diluted 1:5 with saline and assayed again. Multiply the results by 5.

QUALITY SYSTEM CERTIFIED
ISO 9001 ISO 13485

Glory Diagnostics
Manufactured in the Spain

Urine

Dilute the sample 1:50 with distilled water and multiply the result by 50.

If results are to be expressed as SI units apply:
 $\text{mg/dL} \times 0.1665 = \text{mmol/L}$

To convert urea mass units to those of urea nitrogen apply:
 $\text{mg/dL} \times 0.467 = \text{mg/dL BUN}$

REFERENCE VALUES⁵

Serum, plasma

Newborns (< 10 days)	6.4 - 53.5 mg/dL (1.1 - 9.0 mmol/L)
Adults (12-60 years)	15 - 40 mg/dL (2.5 - 6.6 mmol/L)

In adults over 60 years of age, the reference interval is 17-50 mg/dL (2.8-8.3 mmol/L) and the concentrations tend to be slightly higher in males than in females.

Urine

Adults (normal diet)	26 - 43 g/24-h (428 - 714 mmol/24-h)
----------------------	--------------------------------------

A high-protein diet causes significant increases in plasma urea concentrations and urinary excretion. It is recommended that each laboratory establishes its own reference range.

QUALITY CONTROL

The use of a standard to calculate results allows to obtain an accuracy independent of the system or instrument used. To ensure adequate quality control (QC), each run should include a set of controls (normal and abnormal) with assayed values handled as unknowns.

REF BC600 HUMAN MULTISERA NORMAL
 Borderline level of urea. Assayed.

REF BC650 HUMAN MULTISERA ABNORMAL
 Elevated level of urea. Assayed.

If the values are found outside of the defined range, check the instrument, reagents and procedure. Each laboratory should establish its own Quality Control scheme and corrective actions if controls do not meet the acceptable tolerances.

CLINICAL SIGNIFICANCE

Urea is the chief end product of protein metabolism in the body. The importance of the urea concentration in blood lies in its value as an indicator of kidney function.

Azotemia (an abnormal increase in plasma urea level) is seen mainly in renal disorders, dehydration, increase protein catabolism, high-protein diets, or gastrointestinal hemorrhage. There are two types of azotemia. The first, *prerenal azotemia*, is caused by impaired perfusion of the kidneys due to decreased cardiac output or for any of the former causes. The second, *postrenal azotemia*, is caused by an obstruction in the urine outflow such as nephrolithiasis, prostatism, and tumors of the genitourinary tract.

The clinical significance of the urea level in plasma is usually determined in conjugation with the plasma creatinine level. In *prerenal azotemia*, an increase in the plasma urea level is usually associated with a normal plasma creatinine level, where as in *postrenal azotemia*, there is an increase in both the urea and the plasma creatinine levels. A decrease in the urea plasma level may be associated with acute dehydration, malnutrition, and pregnancy.

NOTES

- Collect a 24-hour urine specimen into a plastic bottle free of preservatives. Keep the sample refrigerated to minimize urea hydrolysis by microorganisms or other agents.
- This method may be used with different instruments. Any application to an instrument should be validated to demonstrate that results meets the performance characteristics of the method. It is recommended to validate periodically the instrument. Contact to the distributor for any question on the application method.
- Clinical diagnosis should not be made on findings of a single test result, but should integrate both clinical and laboratory data.

ANALYTICAL PERFORMANCE

- **Detection Limit** : 4.79 mg/dL

- **Linearity** : Up to 300 mg/dL

- **Precision**:

mg/dL	Within-run		Between-run	
Mean	62.3	141.06	62.3	141.06
SD	2.18	5.76	2.91	5.86
CV%	3.33	4.28	4.68	4.16
N	10	10	10	10

- **Sensitivity** : 8.900 mA/min / mg/dL Urea.

- **Correlation**. This assay (y) was compared with a similar commercial method (x). The results were:

$$N = 50 \quad r = 0.99 \quad y = 0.923x + 0.4987$$

The analytical performances have been generated using on automatic instrument. Results may vary depending on the instrument.

REFERENCES

1. Chaney, A.L., and Marbach, E.P. Clin. Chem. 8 : 132 (1962).
2. Searcy, R.L., Reardon, J.E., and Foreman, J.A. Am. J. Clin. Technol. 33 : 15-20 (1967).
3. Young D.S. Effects of drugs on clinical laboratory tests, 5th ed. AACC Press, 2000.
4. Patton, C.S., and Crouch, S.R. Anal. Chem. 49 : 464 (1977)
5. Tietz. N.W. Clinical Guide to Laboratory Tests, 3rd Edition. W.B. Saunders Co. Philadelphia, PA. (1995).
6. Friedman and Young. Effects of disease on clinical laboratory tests. 5th ed. AACC (Press 2000).

QUALITY SYSTEM CERTIFIED
 ISO 9001 ISO 13485

 **Glory Diagnostics**
 Manufactured in the Spain

Lampiran 8. Kontrol Suhu Penyimpanan

No	Suhu (°C)
1	20,3
2	20,7
3	20,2
4	20,0
5	20,0
6	21,1
7	20,0
8	19,8
9	20,2
10	20,1
11	20,0
12	20,0
13	20,2
14	20,0
15	20,3
16	20,0
17	19,8
18	20,6
19	20,1
20	20,2
21	20,0
22	19,2
23	20,3
24	20,1
25	20,0
26	19,8
27	20,3
28	20,0
29	20,2
30	20,0
31	19,6
32	19,3
33	20,1
34	21,2
35	19,7
36	20,0
37	19,5
38	20,3
39	20,5
40	21,2


Lampiran 9. Quality Control

A. Periode Pendahuluan

PERIODE PENDAHULUAN UREUM		RANGE KONTROL
Nomor	Hasil Pemeriksaan Serum Kontrol Periode Pendahuluan	
1	43,38	36,2 - 49
2	43,79	
3	43,55	
4	44,24	
5	43,19	
6	44,92	
7	44,89	
8	45,17	
9	41,33	
10	40,25	
11	40,98	
12	41,05	
13	41,92	
14	43,25	
15	42,50	
16	43,75	
17	43,53	
18	44,16	
19	44,51	
20	44,33	
Mean	43,23	Rerata
SD	1,44	
Mean + 1SD	44,68	
Mean + 2SD	46,12	Batas Peringatan Atas
Mean + 3SD	47,56	Batas Kontrol Atas
Mean - 1SD	41,79	
Mean - 2SD	40,35	Batas Peringatan Bawah
Mean - 3SD	38,91	Batas Kontrol Bawah

B. Periode Kontrol

Periode Kontrol	
Hari Perlakuan	Hasil Pemeriksaan Serum kontrol periode kontrol
Hari 0	44,24
Hari 20	44,51
Hari 30	41,92
Hari 40	42,16



Lampiran 10. Hasil Analisis Deskriptif

PERLAKUKAN	Hari 0	Hari 20	Hari 30	Hari 40
kadar Ureum	43,38	43,17	42,33	41,48
	43,79	43,23	41,73	42,05
	43,55	43,75	42,90	41,08
	44,24	43,53	42,37	41,33
	43,19	44,16	42,27	40,25
	44,92	44,51	43,22	40,98
	44,89	43,21	41,92	41,05
	45,17	42,71	43,25	42,16
RATA-RATA	44,14	43,53	42,50	41,30
Minimum	43,19	42,71	41,73	40,25
Maksimum	45,17	44,16	43,22	42,16
Standar diviasi	0,77	0,58	0,56	0,61
Coefisien Variarant	1,74	1,33	1,32	1,48
Nilai Bahan Kontrol		42,6		
Bias	3,62	2,19	0,24	3,06
PERLAKUKAN	0 hari dan 20 hari	0 hari dan 30 hari	0 hari dan 40 hari	
Selisihh				
Pesentasi Selisihh	1,38%	3,72%	6,44%	

Lampiran 11. Hasil Uji Statistika

1. Uji Distribusi data Kadar Ureum terhadap serum kontrol *liofilisat* komersial setelah rekonstitusi yang disimpan pada suhu -20°C selama 0 hari, 20 hari, 30 hari dan 40 hari.

Hipotesis

H_0 : Data Berdistribusi Normal

H_a : Data tidak berdistribusi normal

Ketentuan

H_0 diterima jika $Sig (Shapiro-Wilk) \geq 0,05$

H_0 ditolak jika $Sig (Shapiro-Wilk) < 0,05$

Hasil

Tests of Normality						
	Kolmogorov-Smirnov ^a			Shapiro-Wilk		
	Statistic	df	Sig.	Statistic	df	Sig.
Standardized Residual for Hari0	,208	8	,200*	,903	8	,304
Standardized Residual for Hari20	,198	8	,200*	,955	8	,759
Standardized Residual for Hari30	,214	8	,200*	,918	8	,416
Standardized Residual for Hari40	,178	8	,200*	,946	8	,669
*. This is a lower bound of the true significance.						
a. Lilliefors Significance Correction						

Keputusan

Data Hasil Kadar Ureum terhadap serum kontrol *liofilisat* komersial setelah rekonstitusi yang disimpan pada suhu -20°C selama 0 hari, 20 hari, 30 hari dan 40 hari karena pada Sig pada *Shapiro-Wilk* $\geq 0,05$.

Kesimpulan

Data Hasil Kadar Ureum terhadap serum kontrol *liofilisat* komersial setelah rekonstitusi yang disimpan pada suhu -20°C selama 0 hari, 20 hari, 30 hari dan 40 hari berdistribusi normal

2. Uji Homogenitas Data (Kesamaan Varians)

Hipotesis

H_0 : Data homogen (varian sama)

H_a : Data tidak homogen (varian tidak sama)

Ketentuan

H_0 diterima jika Sig (*Mauchly's Test of Sphericity*) $\geq 0,05$

H_0 ditolak jika Sig (*Mauchly's Test of Sphericity*) $< 0,05$

Hasil

Mauchly's Test of Sphericity ^a							
Measure: MEASURE_1							
Within Subjects Effect	Mauchly's W	Approx. Chi-Square	df	Sig.	Epsilon ^b		
					Greenhouse-Geisser	Huynh-Feldt	Lower-bound
Hari	,585	3,070	5	,693	,745	1,000	,333
Tests the null hypothesis that the error covariance matrix of the orthonormalized transformed dependent variables is proportional to an identity matrix.							
a. Design: Intercept							
Within Subjects Design: Hari							
b. May be used to adjust the degrees of freedom for the averaged tests of significance. Corrected tests are displayed in the Tests of Within-Subjects Effects table.							

Keputusan

H_0 diterima karena Sig (0,693) $\geq 0,05$

Kesimpulan

Data homogen, maka untuk membuat keputusan uji *Repeated Measures Anova* sig yang dibaca pada *Spherecity Assumed*.

3. Uji Beda Lebih Dua Sampel Berpasangan (*Repeated Measures Anova*)

Hipotesis

H_0 : Tidak ada pengaruh penyimpanan serum kontrol komersial setelah rekonstitusi pada suhu -20°C terhadap hasil kadar Ureum yang disimpan pada 0 hari, 20 hari, 30 hari dan 40 hari

H_a : Ada pengaruh penyimpanan serum kontrol komersial setelah rekonstitusi pada suhu -20°C terhadap hasil kadar Ureum yang disimpan pada 0 hari, 20 hari, 30 hari dan 40 hari

Ketentuan

H_0 diterima jika $Sig \geq 0,05$

H_0 ditolak jika $Sig < 0,05$

Hasil

Tests of Within-Subjects Effects

Measure: MEASURE_1

Source		Type III Sum of Squares	df	Mean Square	F	Sig.
Hari	Sphericity Assumed	37,338	3	12,446	31,082	,000
	Greenhouse-Geisser	37,338	2,235	16,709	31,082	,000
	Huynh-Feldt	37,338	3,000	12,446	31,082	,000
	Lower-bound	37,338	1,000	37,338	31,082	,001
Error(Hari)	Sphericity Assumed	8,409	21	,400		
	Greenhouse-Geisser	8,409	15,642	,538		
	Huynh-Feldt	8,409	21,000	,400		
	Lower-bound	8,409	7,000	1,201		

Keputusan

H_0 Ditolak karena pada *Sphericity Assumed* $(0,000) < 0,05$

Kesimpulan

ada pengaruh penyimpanan serum kontrol komersial setelah rekonstitusi pada suhu -20°C terhadap hasil kadar Ureum yang disimpan pada 0 hari, 20 hari, 30 hari dan 40 hari.

4. Besar Pengaruh Variabel Bebas Terhadap Variabel Terikat

Pairwise Comparisons

Measure: Ureum

(I) Hari	(J) Hari	Mean Difference (I-J)	Std. Error	95% Confidence Interval for Difference ^b		
				Sig. ^b	Lower Bound	Upper Bound
1	2	,608	,376	,903	-,761	1,976
	3	1,642*	,265	,003	,680	2,605

	4	2,844*	,283	,000	1,815	3,873
2	1	-,608	,376	,903	-1,976	,761
	3	1,035*	,255	,029	,107	1,963
	4	2,236*	,399	,005	,785	3,688
3	1	-1,642*	,265	,003	-2,605	-,680
	2	-1,035*	,255	,029	-1,963	-,107
	4	1,201*	,290	,026	,146	2,257
4	1	-2,844*	,283	,000	-3,873	-1,815
	2	-2,236*	,399	,005	-3,688	-,785
	3	-1,201*	,290	,026	-2,257	-,146

Based on estimated marginal means

*. The mean difference is significant at the ,05 level.

b. Adjustment for multiple comparisons: Bonferroni.

Jika sig lebih >0,005 maka tidak ada penurunan

Maka kadar ureum pada serum kontrol *liofilisat* komersial setelah rekonstitusi dan disimpan selama 20 hari pada suhu -20°C adalah tidak mengalami penurunan

Rata-rata penurunan kadar ureum pada serum kontrol *liofilisat* komersial setelah rekonstitusi dan disimpan selama 30 hari pada suhu -20°C adalah 1,64 mg/dL (CI 95% = 0,680 - 2,605)

Rata-rata penurunan kadar ureum pada serum kontrol *liofilisat* komersial setelah rekonstitusi dan disimpan selama 40 hari pada suhu -20°C adalah 2,84 mg/dL (CI 95% = 1,815 - 2,605)

Lampiran 12. Dokumentasi Penelitian



Rekonstitusi Serum Kontrol



Pemisahan Serum Kontrol Ke Cup Serum



Reagen Ureum



Pemantauan Serum Control



Pembuatan Reagen Kerja



Pemeriksaan Kadar Ureum